

TEST REPORT

Doc. No : GTL/D/7.5/02

ULR-TC596126700006967F

Page 1 of 1

Report Issued To:	Sterling Accuris Wellness Pvt.Ltd. Sankalp Square II, Old Sharda Mandir Cross roads B/s. Jalaram Mandir railway Crossing, Ellisbridge,, Ahmedabad, Gujarat Phone No : 9726753939 Email : report.accuris@gmail.com	Test Report No	: GTL/03260729007
		Date of Receipt	: 28/07/2026
		Date of Issue	: 29/07/2026
		Customer's Ref. No	: N.M

Sample Described as	: Water	Mfg. Date	: N.M.
Sample Qty	: 500 ml	Exp. Date	: N.M.
Packing Mode	: Sample Packed in Plastic Bags	B.No	: N.M
Sample Condition	: Satisfactory		
Marking	Lab ID : 072600118285, Specimen ID : 97072600118285, Collection Date : 27-Jul-2026(16:56), Patient Name : Dialysis Department 7th Floor Part 1 (U/)		
Sample Drawn By	: Customer		
Date of Starting of Test	: 28/07/2026	Date of Completion	: 29/07/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

K. M. Amin
Krishna Amin
Autho. Signatory

*** END OF REPORT ***



TC-5961

TEST REPORT

Doc. No : GTL/D/7.5/02

ULR-TC596126700006968F

Page 1 of 1

Report Issued To: Sterling Accuris Wellness Pvt.Ltd. Sankalp Square II, Old Sharda Mandir Cross roads B/s.Jalaram Mandir railway Crossing, Ellisbridge,, Ahmedabad, Gujarat Phone No : 9726753939 Email : report.accuris@gmail.com	Test Report No : GTL/03260729008 Date of Receipt : 28/07/2026 Date of Issue : 29/07/2026 Customer's Ref. No : N.M
---	--

Sample Described as : Water Sample Qty : 500 ml Packing Mode : Sample Packed in Plastic Bags Sample Condition : Satisfactory Marking : Lab ID : 072613401882, Specimen ID : 97072613401882, Collection Date : 27-Jul-2026(12:31), Patient Name : Dialysis Department (U/) Sample Drawn By : Customer Date of Starting of Test : 28/07/2026	Mfg. Date : N.M. Exp. Date : N.M. B.No : N.M. Date of Completion : 29/07/2026
--	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

g.n.m
Krishna Amin
Autho. Signatory

*** END OF REPORT ***

TEST REPORT

Doc. No : GTL/D/7.5/02

ULR-TC596126700006969F

Page 1 of 1

Report Issued To:	Sterling Accuris Wellness Pvt.Ltd. Sankalp Square II, Old Sharda Mandir Cross roads B/s. Jalaram Mandir railway Crossing, Ellisbridge,, Ahmedabad, Gujarat Phone No : 9726753939 Email : report.accuris@gmail.com	Test Report No	: GTL/03260729009
		Date of Receipt	: 28/07/2026
		Date of Issue	: 29/07/2026
		Customer's Ref. No	: N.M

Sample Described as	: Water	Mfg. Date	: N.M.
Sample Qty	: 500 ml	Exp. Date	: N.M.
Packing Mode	: Sample Packed in Plastic Bags	B.No	: N.M
Sample Condition	: Satisfactory		
Marking	: Lab ID : 072652400580, Specimen ID : 97072652400580, Collection Date : 25-Jul-2026(19:49), Patient Name : Mr.Dialysis R.O Unit Water-Sos Hospital Mahuva (M/)		
Sample Drawn By	: Customer		
Date of Starting of Test	: 28/07/2026	Date of Completion	: 29/07/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

8.0.10
Krishna Amin
Autho. Signatory

*** END OF REPORT ***