

**TEST REPORT**

Doc. No : GTL/D/7.5/02

ULR-TC596126700006984F

Page 1 of 1

Report Issued To: Krishna Shalby Hospital Bopal-Ghna road Bopal,Ahmedabad,, Ahmedabad, Gujarat Phone No : 9574237056 Email : icd.krishna@shalby.in	Test Report No : GTL/03260729028 Date of Receipt : 27/07/2026 Date of Issue : 29/07/2026 Customer's Ref. No : N.M
---	--

Sample Described as : Water Sample Qty : 2 Ltr Packing Mode : Sample Packed in Plastic Bottle Sample Condition : Satisfactory Marking : Water Clture Sample (CT-OT) Sampling Location : Krishna Shalby Hospital Sample Drawn By : Mr. Ketul Pathiyar (Representative of Gujarat Testlab Private Limited) Date of Starting of Test : 27/07/2026	Sampling Plan : as per GTL/GP/SOP/046 Sampling Date : 27.07.2026 Environ. Condition : Ambient Date of Completion : 29/07/2026
---	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

S.M.Amin
Krishna Amin
Autho. Signatory

*** END OF REPORT ***