



TC-5961

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

ULR-TC596126700006985F

Report Issued To: **Shalby Hospital (Naroda)..**
S.No:679, Nr. Haridarsan Cross Road,
Naroda,Ahmedabad., Ahmedabad,
Gujarat
Phone No : 9512049805
Email : quality.naroda@shalby.org

Test Report No : GTL/03260729029
Date of Receipt : 27/07/2026
Date of Issue : 29/07/2026
Customer's Ref. No : N.M

Sample Described as : **Water**
Sample Qty : 2 Liter
Packing Mode : Sample Packed in Plastic Bottle
Sample Condition : Satisfactory
Marking : Dialysis Water R.O Tank-2
Sampling Plan : as per GTL/GP/SOP/046
Sampling Date : 27.07.2026
Environ. Condition : Ambient
Sampling Location : Shalby Hospital (Naroda)..
Sample Drawn By : Mr.Mayank solanki (Representative of Gujarat Testlab Pvt Ltd)
Date of Starting of Test : 27/07/2026
Date of Completion : 29/07/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Up	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul

Ketul Padhiyar
Reviewed By



For Verification

*** END OF REPORT ***

FOR, GUJARAT TESTLAB PVT. LTD

S.M.Amin

Krishna Amin
Autho. Signatory