

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1


Report Issued To : Nexcellife Research Labs. Opp. Jalaram Kanaya Chhatralay Altius -3 Road, Thaltej-380059, Ahmedabad, Gujarat Phone No : 7600411219 Email : Parth.Divan@nexcellife.com	Test Report No : GTL/03260730023/NS Date of Receipt : 27/07/2026 Date of Issue : 30/07/2026 Customer's Ref. No : N.M
---	--

Sample Described as : Water.	Mfg. Date : N.M.
Sample Qty : 1 Ltr	Exp. Date : N.M.
Packing Mode : Sample Packed in Plastic Tube	B.No : N.M
Sample Condition : Satisfactory	
Marking : Equipment ID : WPS-01	
Sample Drawn By : Customer	
Date of Starting of Test : 27/07/2026	Date of Completion : 30/07/2026

Sr No	Quality Characteristics	Result	Test Method
Chemical Water			
1	Conductivity at 25 C us/cm	0.056	APHA 24 th Edition
Biological Water			
1	Total Bacterial Count cfu/ml	<1 cfu	IS:5402 (P-1): 2021

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Mamta Bhavsar
Reviewed By



For Verification


Hemal Darji-Trupti Thakor.
Autho. Signatory

*** END OF REPORT ***