

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

ULR-TC596126700007014F

Report Issued To: Apex Kidney Care (Gujarat Kidney Hospital Vadodra)
Vadodra, Vadodara, Gujarat
Phone No : 7016064148
Email : mitesh.machhi@apexkidneycare.in

Test Report No : GTL/03260730032
Date of Receipt : 25/07/2026
Date of Issue : 30/07/2026
Customer's Ref. No : N.M

Sample Described as : Water
Sample Qty : 2.0 Ltr
Packing Mode : Sample Packed in Plastic Bottle
Sample Condition : Satisfactory
Marking : R.O. Water
Mfg. Date : N.M.
Exp. Date : N.M.
B.No : N.M.

Sample Drawn By : Customer
Date of Starting of Test : 25/07/2026

Date of Completion : 30/07/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

8.10.10
Krishna Amin
Autho. Signatory

*** END OF REPORT ***

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To : Apex Kidney Care (Gujarat Kidney Hospital Vadodra) Vadodra , Vadodara, Gujarat Phone No : 7016064148 Email : mitesh.machhi@apexkidneycare.in	Test Report No : GTL/03260730032/NS Date of Receipt : 25/07/2026 Date of Issue : 30/07/2026 Customer's Ref. No : N.M
---	---

Sample Described as : Water Sample Qty : 2.0 Ltr Packing Mode : Sample Packed in Plastic Bottle Sample Condition : Satisfactory Marking : R.O.Water	Mfg. Date : N.M. Exp. Date : N.M. B.No : N.M.
Sample Drawn By : Customer Date of Starting of Test : 25/07/2026	Date of Completion : 30/07/2026

Sr No	Quality Characteristics	Result	Test Method
Biological Water			
1	Total Bacterial Count cfu/ml	08 cfu	IS: 5402 (P-1)2021

Note : NM = Not Mentioned, Customer requested for the above test only

Ketul Padhiyar

Reviewed By



For Verification

*** END OF REPORT ***

FOR, GUJARAT TESTLAB PVT. LTD

Krishna Amin

Autho. Signatory