



ULR-TC596126700007038F

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To: Star Therapeutic Pvt. Ltd. (Star Hospital)
Shed NO : 2 & 3, Arvind Estate Indica
Laboratory, Bapunagar, Ahmedabad,
Gujarat
Phone No : 9598034293
Email : startherapeutic@gmail.com

Test Report No : GTL/03260731003
Date of Receipt : 27/07/2026
Date of Issue : 31/07/2026
Customer's Ref. No : N.M

Sample Described as : Water
Sample Qty : 2 Ltr
Packing Mode : Sample Packed in Plastic Bottle
Sample Condition : Satisfactory
Marking : Dialysis Water Tank

Sampling Plan : as per GTL/GP/SOP/046
Sampling Date : 27.07.2026
Environ. Condition : Ambient

Sampling Location : Star Therapeutic Pvt. Ltd. (Star Hospital)
Sample Drawn By : Mr. Satish Rana (Representative of Gujarat Testlab Private Limited)
Date of Starting of Test : 27/07/2026

Date of Completion : 31/07/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

CONFORM

Remark: In respect of the test performed the submitted sample to the above Specification for the tested parameters.

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

*** END OF REPORT ***

FOR, GUJARAT TESTLAB PVT. LTD

Krishna Amin
Krishna Amin
Autho. Signatory



TC-5961

ULR-TC596126700007039F

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To: Star Therapeutic Pvt. Ltd. (Star Hospital)
Shed NO : 2 & 3, Arvind Estate Indica
Laboratory, Bapunagar, Ahmedabad,
Gujarat
Phone No : 9598034293
Email : startherapeutic@gmail.com

Test Report No : GTL/03260731004
Date of Receipt : 27/07/2026
Date of Issue : 31/07/2026
Customer's Ref. No : N.M

Sample Described as : Water

Sample Qty : 2 Ltr

Packing Mode : Sample Packed in Plastic Bottle

Sample Condition : Satisfactory

Marking : Dialysis Water

Sampling Plan : as per GTL/GP/SOP/046

Sampling Date : 27.07.2026

Environ. Condition : Ambient

Sampling Location : Star Therapeutic Pvt. Ltd. (Star Hospital)

Sample Drawn By : Mr. Satish Rana (Representative of Gujarat Testlab Private Limited)

Date of Starting of Test : 27/07/2026

Date of Completion : 31/07/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above Specification for the tested parameters.

Note : NM = Not Mentioned, Customer requested for the above test only

CONFORM



For NABL Scope

Ketui
Ketui Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

S. Krishna
Krishna Amin
Autho. Signatory

*** END OF REPORT ***