

**TEST REPORT**

Doc. No : GTL/D/7.5/02

ULR-TC596126700007118F

Page 1 of 1

Report Issued To:	Sterling Accuris Wellness Pvt. Ltd. Sankalp Square II, Old Sharda Mandir Cross roads B.S Jalaram Mandir railway Crossing, Ellisbridge,, Ahmedabad, Gujarat Phone No : 9726753939 Email : report.accuris@gmail.com	Test Report No	: GTL/03260803014
		Date of Receipt	: 31/07/2026
		Date of Issue	: 03/08/2026
		Customer's Ref. No	: N.M

Sample Described as	: Water	Mfg. Date	: N.M.
Sample Qty	: 50 ml	Exp. Date	: N.M.
Packing Mode	: Sample Packed in Plastic Bottle	B.No	: N.M
Sample Condition	: Satisfactory		
Marking	: Lab ID : 072671600418, Specimen Type : 97072671600418, Collection Date : 30-Jul-2026(15:34), Patient Name : Water C/S (U/)		
Sample Drawn By	: Customer		
Date of Starting of Test	: 31/07/2026	Date of Completion	: 03/08/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological				
Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the parameters tested

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD

For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

Do.A.Mu
Krishna Amin
Autho. Signatory

*** END OF REPORT ***