

TEST REPORT

Doc. No : GTL/D/7.5/02

ULR-TC596126700007170F

Page 1 of 1

Report Issued To: Epic Multispecialty Hospital Opp.Pandit Dindayal Upadhyay Hall, Behind Rajpath Club, S.G.Highway Bodakdev, Ahmedabad, Gujarat Phone No : Email : biomed@epichospital.com	Test Report No : GTL/03260804013 Date of Receipt : 30/07/2026 Date of Issue : 04/08/2026 Customer's Ref. No : N.M
--	--

Sample Described as : Water Sample Qty : 2 Ltr Packing Mode : Sample Packed in Plastic Bottle Sample Condition : Satisfactory Marking : Dialysis Water Port No : 01 Sampling Location : Epic Multispecialty Hospital Sample Drawn By : Mr. Satish Rana (Representative of Gujarat Testlab Private Limited) Date of Starting of Test : 30/07/2026	Sampling Plan : as per GTL/GP/SOP/055 Sampling Date : 30.07.2026 Environ. Condition : Ambient Date of Completion : 04/08/2026
---	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the parameters tested

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

Bohina
Krishna Amin
Autho. Signatory

*** END OF REPORT ***

TEST REPORT

Doc. No : GTL/D/7.5/02

ULR-TC596126700007171F

Page 1 of 1

Report Issued To: Epic Multispecialty Hospital Opp.Pandit Dindayal Upadhyay Hall, Behind Rajpath Club, S.G.Highway Bodakdev, Ahmedabad, Gujarat Phone No : Email : biomed@epichospital.com	Test Report No : GTL/03260804014 Date of Receipt : 30/07/2026 Date of Issue : 04/08/2026 Customer's Ref. No : N.M
--	--

Sample Described as : Water Sample Qty : 2 Ltr Packing Mode : Sample Packed in Plastic Bottle Sample Condition : Satisfactory Marking : Dialysis Water Port No : 07 Sampling Location : Epic Multispecialty Hospital Sample Drawn By : Mr. Satish Rana (Representative of Gujarat Testlab Private Limited) Date of Starting of Test : 30/07/2026	Sampling Plan : as per GTL/GP/SOP/055 Sampling Date : 30.07.2026 Environ. Condition : Ambient Date of Completion : 04/08/2026
---	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the parameters tested

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

Krishna Amin
Krishna Amin
Autho. Signatory

*** END OF REPORT ***

**TEST REPORT**

Doc. No : GTL/D/7.5/02

Page 1 of 1

ULR-TC596126700007172F

Report Issued To: Epic Multispecialty Hospital Opp.Pandit Dindayal Upadhyay Hall, Behind Rajpath Club, S.G.Highway Bodakdev, Ahmedabad, Gujarat Phone No : Email : biomed@epichospital.com	Test Report No : GTL/03260804015 Date of Receipt : 30/07/2026 Date of Issue : 04/08/2026 Customer's Ref. No : N.M
--	--

Sample Described as : Water Sample Qty : 2 Ltr Packing Mode : Sample Packed in Plastic Bottle Sample Condition : Satisfactory Marking : Dialysis Water Port No : 19 Sampling Location : Epic Multispecialty Hospital Sample Drawn By : Mr. Satish Rana (Representative of Gujarat Testlab Private Limited) Date of Starting of Test : 30/07/2026	Sampling Plan : as per GTL/GP/SOP/055 Sampling Date : 30.07.2026 Environ. Condition : Ambient Date of Completion : 04/08/2026
---	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the parameters tested

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

Krishna Amin
Krishna Amin
Autho. Signatory

*** END OF REPORT ***

**TEST REPORT**

Doc. No : GTL/D/7.5/02

Page 1 of 1

ULR-TC596126700007173F

Report Issued To: Epic Multispecialty Hospital Opp.Pandit Dindayal Upadhyay Hall, Behind Rajpath Club, S.G.Highway Bodakdev, Ahmedabad, Gujarat Phone No : Email : biomed@epichospital.com	Test Report No : GTL/03260804016 Date of Receipt : 30/07/2026 Date of Issue : 04/08/2026 Customer's Ref. No : N.M
--	--

Sample Described as : Water Sample Qty : 2 Ltr Packing Mode : Sample Packed in Plastic Bottle Sample Condition : Satisfactory Marking : Dialysis Water Port No : 12 Sampling Location : Epic Multispecialty Hospital Sample Drawn By : Mr. Satish Rana (Representative of Gujarat Testlab Private Limited) Date of Starting of Test : 30/07/2026	Sampling Plan : as per GTL/GP/SOP/055 Sampling Date : 30.07.2026 Environ. Condition : Ambient Date of Completion : 04/08/2026
---	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the parameters tested

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

S.M.N.
Krishna Amin
Autho. Signatory

*** END OF REPORT ***