

TEST REPORT

Doc. No : GTL/D/7.5/02

ULR-TC596126700007190F

Page 1 of 1

Report Issued To: Unipath Specialty Laboratory Ltd. 7 Pasi House, Beside Sahjanand College, Opp. Kamdhenu Complex, Panjrapole, Amdawadi, Ahmedabad, Gujarat Phone No : Email : info@unipath.in	Test Report No : GTL/03260804033 Date of Receipt : 03/08/2026 Date of Issue : 04/08/2026 Customer's Ref. No : N.M
---	--

Sample Described as : Water Sample Qty : 50 ml Packing Mode : Sample Packed in Sterile Container Sample Condition : Satisfactory Marking : Lab ID : 082615200026, Specimen Id : 97082615200026, Collection Date : 01-Aug-2026 (11:58), Patient Name : Dialysis Unit Metas (U/)	Mfg. Date : N.M. Exp. Date : N.M. B.No : N.M.
Sample Drawn By : Customer Date of Starting of Test : 03/08/2026	Date of Completion : 04/08/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per USP	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above Specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketul
Ketul Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

Krishna Amin
Krishna Amin
Autho. Signatory

*** END OF REPORT ***