

TEST REPORT

Doc. No : GTL/D/7.5/02

ULR-TC596126700007235F

Page 1 of 1

Report Issued To: SGVP Holistic Hospital SGVP Campus, S.G Highway, Chharodi,, Ahmedabad, Gujarat Phone No : Email : chirag.parmar@sgvpvh.org	Test Report No : GTL/03260806006 Date of Receipt : 01/08/2026 Date of Issue : 06/08/2026 Customer's Ref. No : N.M
---	--

Sample Described as Sample Qty Packing Mode Sample Condition Marking Sampling Location Sample Drawn By Date of Starting of Test	: Water : 2 Ltr : Sample Packed in Plastic Bottle : Satisfactory : Dialysis R.O Water, Location : Old Dialysis Unit Bed no.02 : SGVP Holistic Hospital : Mr. Satish Dataniya (Representative of Gujarat Testlab Private Limited) : 01/08/2026	Sampling Plan : as per GTL/GP/SOP/055 Sampling Date : 01.08.2026 Environ. Condition : Ambient Date of Completion : 06/08/2026
--	--	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only



For NABL Scope

Ketun
Ketun Padhiyar
Reviewed By



For Verification

FOR, GUJARAT TESTLAB PVT. LTD

Shravan
Krishna Amin
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.