

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

ULR-TC596126700017621F

Report Issued To: Krishna Shalby Hospital.
Bopal-Ghuma Road, Bopal,
Ahmedabad, Gujarat
Phone No : 9574237056
Email : amaccounts.krishna@shalby

Test Report No : GTL/03260818005
Date of Receipt : 13/08/2026
Date of Issue : 18/08/2026
Customer's Ref. No : N.M

Sample Described as : Water
Sample Qty : 1 Ltr
Packing Mode : Sample Packed in Plastic Bottle
Sample Condition : Satisfactory
Marking : Dialysis Water
Sampling Plan : as per GTL/GP/SOP/055
Sampling Date : 13.08.2026
Environ. Condition : Ambient
Sampling Location : Krishna Shalby Hospital.
Sample Drawn By : Mr.Satish Rana (Representative of Gujarat Testlab Private Limited)
Date of Starting of Test : 13/08/2026
Date of Completion : 18/08/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	<0.25 EU/ml	As per Usp	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the tested parameters.

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD



For NABL Scope

Ketul Padhiyar
Reviewed By



For Verification

Krishna Amin
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.