

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

ULR-TC596126700019036F

Report Issued To:	Sterling Accuris Wellness Pvt. Ltd. Sankalp Square II, Old Sharda Mandir Cross roads B.S Jalaram Mandir railway Crossing, Ellisbridge,, Ahmedabad, Gujarat Phone No : 9726753939 Email : report.accuris@gmail.com	Test Report No	: GTL/03260918020
		Date of Receipt	: 14/09/2026
		Date of Issue	: 18/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Water	Mfg. Date	: N.M.
Sample Qty	: 500 ml	Exp. Date	: N.M.
Packing Mode	: Sample Packed in Plastic Container	B.No	: N.M
Sample Condition	: Satisfactory		
Marking	: Lab ID : 092600110550, Specimen Id : 97092600110550, Collection Date : 14-Sep-2026 (13:36), Patient Name : Dialysis Department 2nd Floor (U/)		
Sample Drawn By	: Customer	Date of Completion	: 18/09/2026
Date of Starting of Test	: 14/09/2026		

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Biological Water				
1	Bacterial Endotoxins Test (Lal Test)	< 0.25 EU/ml	As per USP	Max 2.0 EU/ml

Remark: In respect of the test performed the submitted sample to the above specification for the parameters tested

CONFORM

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD



For NABL Scope

Ketul
Ketul Padhiyar

Reviewed By



For Verification

g.o.m.
Krishna Amin

Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.