

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To :	Thermo Fisher Scientific India Pvt Ltd. C/O. Trichem laboratories (Bom.) Pvt. Ltd. Plot No. K-43, MIDC, Ambarnath area, Anandnagar, Jambivili village, Thane, Maharashtra Phone No : Email : raiesh.dinankar@thermofisher.com	Test Report No	: GTL/04260912058/NS
		Date of Receipt	: 11/09/2026
		Date of Issue	: 12/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Buffer Solution	Mfg. Date	: Sep-2026
Sample Qty	: 500 ml	Exp. Date	: N.M.
Packing Mode	: Sample Packed in Plastic Bottle	B.No	: 9674820926
Sample Condition	: Satisfactory		
Marking	: pH buffer 1.68, Bottle No 25 of 60		
Sample Drawn By	: Customer	Date of Completion	: 12/09/2026
Date of Starting of Test	: 11/09/2026		

Sr No	Quality Characteristics	Result	Test Method
Chemical			
Miscellaneous			
1	pH at 25 Deg C	Set-1 : 1.67, Set-2 : 1.69, Set-3 : 1.69	By pH Meter

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Mamta Bhavsar
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

Disclaimer Note

*** END OF REPORT ***

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To : Thermo Fisher Scientific India Pvt Ltd. C/O. Trichem laboratories (Bom.) Pvt. Ltd. Plot No. K-43, MIDC, Ambernath area, Anandnagar, Jambivili village, Thane, Maharashtra Phone No : Email : raiesh.dinankar@thermofisher.com	Test Report No : GTL/04260912059/NS Date of Receipt : 11/09/2026 Date of Issue : 12/09/2026 Customer's Ref. No : N.M
--	---

Sample Described as : Buffer Solution Sample Qty : 500 ml Packing Mode : Sample Packed in Plastic Bottle Sample Condition : Satisfactory Marking : pH buffer 9.18, Bottle No 4 of 60	Mfg. Date : Sep-2026 Exp. Date : N.M. B.No : 9674840926
Sample Drawn By : Customer Date of Starting of Test : 11/09/2026	Date of Completion : 12/09/2026

Sr No	Quality Characteristics	Result	Test Method
Chemical Miscellaneous			
1	pH at 25 Deg C	Set -1 : 9.17, Set-2 : 9.17, Set-3 : 9.18	By pH Meter

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Mamta Bhavsar
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

Disclaimer Note

*** END OF REPORT ***

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To : Thermo Fisher Scientific India Pvt Ltd. C/O. Trichem laboratories (Bom.) Pvt. Ltd. Plot No. K-43, MIDC, Ambarnath area, Anandnagar, Jambivili village, Thane, Maharashtra Phone No : Email : raiesh.dinankar@thermofisher.com	Test Report No : GTL/04260912060/NS Date of Receipt : 11/09/2026 Date of Issue : 12/09/2026 Customer's Ref. No : N.M
---	---

Sample Described as : Buffer Solution	Mfg. Date : Sep-2026
Sample Qty : 500 ml	Exp. Date : N.M.
Packing Mode : Sample Packed in Plastic Bottle	B.No : 9674830926
Sample Condition : Satisfactory	
Marking : pH buffer 7.00, Bottle No 27 of 60	
Sample Drawn By : Customer	
Date of Starting of Test : 11/09/2026	Date of Completion : 12/09/2026

Sr No	Quality Characteristics	Result	Test Method
Chemical			
Miscellaneous			
1	pH at 25 Deg C	Set-1 : 6.99, Set-2 : 7.01, Set-3 : 7.01	By pH Meter

Note : NM = Not Mentioned, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Mamta Bhavsar
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

Disclaimer Note

*** END OF REPORT ***
(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.