

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To:	Valaji Pharma Chem Plot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No	: GTL/04260918010/NS
		Date of Receipt	: 15/09/2026
		Date of Issue	: 18/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Magnesium Bisglycinate 20% IHS	Mfg. Date	: 01.09.2026
Sample Qty	: 30 gm	Exp. Date	: 31.08.2029
Packing Mode	: Sample Packed in Plastic Bags	B.No	: VEI00126
Sample Condition	: Satisfactory		
Marking	: Batch Size : 9000 Kg, Tr No.603		
Sample Drawn By	: Customer		
Date of Starting of Test	: 15/09/2026	Date of Completion	: 18/09/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical				
Industrial and Fine Chemicals				
1	Lead (as Pb) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1
2	Cadmium (as Cd) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.5
3	Arsenic (as As) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1.0
4	Mercury (as Hg) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification,
Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD

Rupesh Desai
Reviewed By



For Verification

Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 014SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 14:50:48
Sample Name G_260915029
Sample Type Sample
Prep Dilution 20.0335
Auto Dilution 1.0000
Total Dilution 20.0335
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	1.282	ppb	Pulse	29.41	0.1020
Cd	111	2	He	1.543	ppb	Pulse	78.44	0.1020
Hg	201	2	He	8.717	ppb	Pulse	500.07	0.1020
Pb	208	2	He	18.342	ppb	Pulse	12592.73	0.1020



TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To:	Valaji Pharma Chem Plot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No	: GTL/04260918011/NS
		Date of Receipt	: 15/09/2026
		Date of Issue	: 18/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Magnesium Malate 15 % IHS	Mfg. Date	: 07.09.2026
Sample Qty	: 30 gm	Exp. Date	: 06.09.2029
Packing Mode	: Sample Packed in Plastic Bags	B.No	: VEI00426
Sample Condition	: Satisfactory		
Marking	: Batch Size : 4050 Kg, Tr No.609		
Sample Drawn By	: Customer		
Date of Starting of Test	: 15/09/2026	Date of Completion	: 18/09/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical				
Industrial and Fine Chemicals				
1	Arsenic (as As) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1
2	Mercury (as Hg) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1
3	Lead (as Pb) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1
4	Cadmium (as Cd) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.5

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification,
Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Rupesh Desai
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 015SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 14:53:51
Sample Name G_260915030
Sample Type Sample
Prep Dilution 19.8998
Auto Dilution 1.0000
Total Dilution 19.8998
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	4.118	ppb	Pulse	22.88	0.1020
Cd	111	2	He	4.905	ppb	Pulse	183.03	0.1020
Hg	201	2	He	0.479	ppb	Pulse	98.05	0.1020
Pb	208	2	He	16.116	ppb	Pulse	5957.68	0.1020



TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To: Valaji Pharma Chem PLot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No : GTL/04260918013/NS Date of Receipt : 15/09/2026 Date of Issue : 18/09/2026 Customer's Ref. No : N.M
---	---

Sample Described as : Magnesium Malate 28 % IHS Sample Qty : 30 gm Packing Mode : Sample Packed in Plastic Bags Sample Condition : Satisfactory Marking : Batch Size : 450 Kg, Tr No.608 Sample Drawn By : Customer Date of Starting of Test : 15/09/2026	Mfg. Date : 07.09.2026 Exp. Date : 06.09.2029 B.No : VEI00326 Date of Completion : 18/09/2026
--	--


Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical Industrial and Fine Chemicals				
1	Arsenic (as As) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1
2	Mercury (as Hg) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1
3	Lead (as Pb) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 3
4	Cadmium (as Cd) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification,
Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Rupesh Desai
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 016SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 14:57:08
Sample Name G_260915031
Sample Type Sample
Prep Dilution 19.4805
Auto Dilution 1.0000
Total Dilution 19.4805
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	3.590	ppb	Pulse	52.29	0.1020
Cd	111	2	He	0.184	ppb	Pulse	13.08	0.1020
Hg	201	2	He	0.515	ppb	Pulse	173.23	0.1020
Pb	208	2	He	1.572	ppb	Pulse	2801.29	0.1020



TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To:	Valaji Pharma Chem Plot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No	: GTL/04260918014/NS
		Date of Receipt	: 15/09/2026
		Date of Issue	: 18/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Calcium Citrate Malate	Mfg. Date	: 31.08.2026
Sample Qty	: 30 gm	Exp. Date	: 30.08.2029
Packing Mode	: Sample Packed in Plastic Bags	B.No	: VEH04526
Sample Condition	: Satisfactory		
Marking	: Batch Size : 500 Kg, Tr No.599		
Sample Drawn By	: Customer		
Date of Starting of Test	: 15/09/2026	Date of Completion	: 18/09/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical				
Industrial and Fine Chemicals				
1	Lead (as Pb) ppm	0.20	By ICP-MS	Max. 3
2	Cadmium (as Cd) ppm	0.21	By ICP-MS	Max. 1
3	Arsenic (as As) ppm	0.072	By ICP-MS	Max. 1
4	Mercury (as Hg) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1
5	Heavy metals mg/kg	< 10	As per USP	Max. 10

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification, Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD

Rupesh Desai
Reviewed By



For Verification

Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 017SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 15:00:15
Sample Name G_260915032
Sample Type Sample
Prep Dilution 19.7512
Auto Dilution 1.0000
Total Dilution 19.7512
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	72.142	ppb	Pulse	2131.13	0.1020
Cd	111	2	He	209.163	ppb	Pulse	18989.59	0.1020
Hg	201	2	He	8.331	ppb	Pulse	4805.42	0.1020
Pb	208	2	He	201.826	ppb	Pulse	140806.39	0.1020



TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To:	Valaji Pharma Chem Plot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No	: GTL/04260918015/NS
		Date of Receipt	: 15/09/2026
		Date of Issue	: 18/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Magnesium Bisglycinate 12.5 %	Mfg. Date	: 25.08.2026
Sample Qty	: 30 gm	Exp. Date	: 24.08.2029
Packing Mode	: Sample Packed in Plastic Bags	B.No	: VEH02926
Sample Condition	: Satisfactory		
Marking	: Batch Size : 2000 Kg,Tr No.583		
Sample Drawn By	: Customer		
Date of Starting of Test	: 15/09/2026	Date of Completion	: 18/09/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical				
Industrial and Fine Chemicals				
1	Arsenic (as As) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1
2	Mercury (as Hg) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1
3	Lead (as Pb) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 3
4	Cadmium (as Cd) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification,
Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD

Rupesh Desai
Reviewed By



For Verification

Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 018SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 15:06:21
Sample Name G_260915033
Sample Type Sample
Prep Dilution 19.7567
Auto Dilution 1.0000
Total Dilution 19.7567
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	4.559	ppb	Pulse	85.56	0.1020
Cd	111	2	He	0.065	ppb	Pulse	6.55	0.1020
Hg	201	2	He	3.012	ppb	Pulse	294.16	0.1020
Pb	208	2	He	1.305	ppb	Pulse	3236.12	0.1020



TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To:	Valaji Pharma Chem Plot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No	: GTL/04260918016/NS
		Date of Receipt	: 15/09/2026
		Date of Issue	: 18/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Magnesium Bisglycinate 12.5 %	Mfg. Date	: 24.08.2026
Sample Qty	: 30 gm	Exp. Date	: 23.08.2029
Packing Mode	: Sample Packed in Plastic Bags	B.No	: VEH02826
Sample Condition	: Satisfactory		
Marking	: Batch Size : 2000 Kg,Tr No.581		
Sample Drawn By	: Customer		
Date of Starting of Test	: 15/09/2026	Date of Completion	: 18/09/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical				
Industrial and Fine Chemicals				
1	Arsenic (as As) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1
2	Mercury (as Hg) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1
3	Lead (as Pb) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 3
4	Cadmium (as Cd) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification,
Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Rupesh Desai
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 019SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 16:09:26
Sample Name G_260915034
Sample Type Sample
Prep Dilution 19.8469
Auto Dilution 1.0000
Total Dilution 19.8469
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	2.019	ppb	Pulse	42.49	0.1020
Cd	111	2	He	0.543	ppb	Pulse	29.42	0.1020
Hg	201	2	He	<0.000	ppb	Pulse	72.84	0.1020
Pb	208	2	He	17.008	ppb	Pulse	11954.81	0.1020



TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To:	Valaji Pharma Chem Plot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No	: GTL/04260918017/NS
		Date of Receipt	: 15/09/2026
		Date of Issue	: 18/09/2026
		Customer's Ref. No	: N.M

Sample Described as	: Magnesium Bisglycinate 20% IHS	Mfg. Date	: 04.09.2026
Sample Qty	: 30 gm	Exp. Date	: 03.09.2029
Packing Mode	: Sample Packed in Plastic Bags	B.No	: VEI00226
Sample Condition	: Satisfactory		
Marking	: Batch Size : 4500 Kg, Tr No.605		
Sample Drawn By	: Customer		
Date of Starting of Test	: 15/09/2026	Date of Completion	: 18/09/2026

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical				
Industrial and Fine Chemicals				
1	Lead (as Pb) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1
2	Cadmium (as Cd) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.5
3	Arsenic (as As) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 1.0
4	Mercury (as Hg) mg/kg	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification,
Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Rupesh Desai
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 020SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 16:12:32
Sample Name G_260915036
Sample Type Sample
Prep Dilution 20.1034
Auto Dilution 1.0000
Total Dilution 20.1034
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	1.482	ppb	Pulse	32.68	0.1020
Cd	111	2	He	0.878	ppb	Pulse	52.30	0.1020
Hg	201	2	He	<0.000	ppb	Pulse	78.75	0.1020
Pb	208	2	He	18.542	ppb	Pulse	13909.41	0.1020

TEST REPORT

Doc. No : GTL/D/7.5/02

Page 1 of 1

Report Issued To: Valaji Pharma Chem Plot No : 475, Post; luna, Ta : Padra , Vadodara, Gujarat Phone No : 8128669330 Email : bhumika@valajipharmachem.co.in	Test Report No : GTL/04260918018/NS Date of Receipt : 15/09/2026 Date of Issue : 18/09/2026 Customer's Ref. No : N.M
---	---

Sample Described as : Calcium Carbonate (Light) BP Sample Qty : 30 gm Packing Mode : Sample Packed in Plastic Bags Sample Condition : Satisfactory Marking : Batch Size : 1000 Kg,Tr No.574 Sample Drawn By : Customer Date of Starting of Test : 15/09/2026	Mfg. Date : 23.08.2026 Exp. Date : 22.08.2029 B.No : VEH02526 Date of Completion : 18/09/2026
---	--

Sr No	Quality Characteristics	Result	Test Method	Customer Specification
Chemical				
Industrial and Fine Chemicals				
1	Arsenic (as As) ppm	0.058	By ICP-MS	Max. 4
2	Mercury (as Hg) ppm	B.L.Q.(Q.L.=0.02)	By ICP-MS	Max. 0.1
3	Lead (as Pb) ppm	0.21	By ICP-MS	Max. 3
4	Cadmium (as Cd) ppm	0.11	By ICP-MS	Max. 1

Remark: In respect of the test performed the submitted sample to the above specification for the Parameter tested

CONFORM

Note : N.M.=Not Mentioned, B.L.Q.=Below Limit of Quantification,
Q.L.=Quantification Limit, Customer requested for the above test only

FOR, GUJARAT TESTLAB PVT. LTD


Rupesh Desai
Reviewed By



For Verification


Hemal Darji
Autho. Signatory

*** END OF REPORT ***

Disclaimer Note

(1) The test results relate only to the tested sample(s) and the applicable parameter(s). Endorsement or certification of the product is neither inferred nor implied. (2) This test report shall not be reproduced, either wholly or in part, and shall not be used as evidence in any court of law or for any advertising or promotional purpose without the prior written approval of Gujarat Testlab Private Limited. (3) Customer requested for the above test only. (4) Unless otherwise agreed with the customer or required by a regulatory authority, tested samples will be disposed of after 21 days from the date of issue of the test report. Perishable samples will be destroyed after testing. (5) Samples not drawn by us unless otherwise stated (6) The total liability of the laboratory is limited to the amount invoiced for the testing services. (7) Test reports are issued based on the information and/or data provided by the customer. (8) The accuracy and validity of the test results may be affected if such information or data is incorrect, incomplete, or misleading. (9) Subject to the Ahmedabad Jurisdiction Only.

Quantitation Report

Data File Name 021SMPL.d
Acq/Data Batch D:\Agilent\ICPMH\1\New folder\16-09-2026.b
Acq Time 16/09/2026 16:15:38
Sample Name G_260915039
Sample Type Sample
Prep Dilution 20.1696
Auto Dilution 1.0000
Total Dilution 20.1696
Acq Mode Spectrum
Last Calib 16/09/2026 16:38:11
Bkg Mode Count Subtraction except for ISTD

External Calibration

FullQuant Table

Name	Mass	Tune Step	Tune Mode	Conc.	Units	Det.	CPS	Integ Time
As	75	2	He	58.029	ppb	Pulse	2526.67	0.1020
Cd	111	2	He	108.336	ppb	Pulse	26094.21	0.1020
Hg	201	2	He	3.229	ppb	Pulse	529.48	0.1020
Pb	208	2	He	208.755	ppb	Pulse	134951.53	0.1020

